

VSL, the National Metrology Institute (NMI) of the Netherlands declares that the Calibration and Measurement Capabilities of the calibration test facilities named in this certificate are granted until the expiration date to:

Emerson Process Management Flow B.V.

Neonstraat 1 | 6718 WX EDE | The Netherlands

During the CMC certification process it was demonstrated that the test facilities named in this certificate apply to the following requirements:

- All instruments in the test facilities are traceable to international accepted measurement standards,
- The calibration models (equations) are correct,
- The Calibration and Measurement Capabilities (CMC) are calculated according the JCGM 100:2008 "Evaluation of measurement data - Guide to the expression of uncertainty in measurement",
- A set of procedures is in place to maintain the CMC for the period of validation.

LF.2007.06.0005

granted on 15 October 2020

This declaration is valid until:
27 October 2022

The CMC certification has been granted for the first time on:
27 June 2007

Test facilities with CMC certification valid under this certificate:

SSF1B, VSL CMC certificate number 39333706, valid until 27 October 2022

SSF2A, VSL CMC certificate number 39333707, valid until 27 October 2022

SSF3A, VSL CMC certificate number 39333708, valid until 27 October 2022


Manager Liquid Flow Metrology VSL, F.M. Smits



V

Thijs
2629
The N
www