



VSL, the National Metrology Institute (NMI) of the Netherlands declares that the Calibration and Measurement Capabilities of the calibration test facilities named in this certificate are granted until the expiration date to:

Thijsseweg 11
2629 JA Delft
The Netherlands
www.vsl.nl

Cameron Measurement Systems Caldon® Ultrasonics Technology Center

1000 McClaren Woods Drive | Coraopolis, PA 15108 | United States of America

During the CMC certification process it was demonstrated that the test facilities named in this certificate apply to the following requirements:

- All instruments in the test facilities are traceable to international accepted measurement standards,
- The calibration models (equations) are correct,
- The Calibration and Measurement Capabilities (CMC) are calculated according the JCGM 100:2008 "Evaluation of measurement data - Guide to the expression of uncertainty in measurement",
- A set of procedures is in place to maintain the CMC for the period of validation.

LF.2008.04.0002

granted on 12 December 2016


This declaration is valid until:
12 December 2018

The CMC certification has been granted for the first time on:

10 April 2008

Test facilities with CMC certification valid under this certificate:

**Large liquid flow meter calibration loop,
VSL CMC certificate number 39332847,
valid until 12 December 2018**



Managing director VSL, Ir. H.C.L. Vos