



VSL, the National Metrology Institute (NMI) of the Netherlands declares that the Calibration and Measurement Capabilities of the calibration test facilities mentioned in this certificate is granted to:

Thijsseweg 11
2629 JA Delft
The Netherlands
www.vsl.nl

Emerson FZE

PO Box 17034 | Jebel Ali Free zone- South (Zone 2), Plot No. S20137 | 17034 Dubai | United Arab Emirates

During the CMC certification process it was demonstrated that the test facilities mentioned in this certificate comply with the following requirements:

- All instruments in the test facilities are traceable to international accepted measurement standards,
- The calibration models (equations) are correct,
- The Calibration and Measurement Capabilities (CMC) are calculated according the JCGM 100:2008 "Evaluation of measurement data - Guide to the expression of uncertainty in measurement",
- A set of procedures is in place to maintain the CMC for the period of validation.

LF.2011.02.0001

granted on 3 July 2020


This CMC Certification is valid until:
16 April 2022

The CMC certification has been granted for the first time on:
7 February 2011

The CMC certification valid for the following test facilities:

TSM1B
VSL CMC certificate number 39333649

TSM2B
VSL CMC certificate number 39333650



Manager Flow Metrology VSL, F.M. Smits