



VSL, the National Metrology Institute (NMI) of the Netherlands declares that the Calibration and Measurement Capabilities of the calibration test facilities mentioned in this certificate is granted to:

Thijsseweg 11
2629 JA Delft
The Netherlands
www.vsl.nl

Stichting Flow Center, Flow Center of Excellence

Leerparkpromenade 50 | 3312KW Dordrecht | The Netherlands

During the CMC certification process it was demonstrated that the test facilities mentioned in this certificate comply with the following requirements:

- All relevant instruments in the test facility are traceable to international accepted measurement standards,
- The calibration models (equations) are correct,
- The Calibration and Measurement Capabilities (CMC) are calculated according the JCGM 100:2008 "Evaluation of measurement data - Guide to the expression of uncertainty in measurement",
- A set of procedures is in place to maintain the CMC for the period of validation.

LF.2018.01.0013


granted on 5 August 2020

This CMC Certification is valid until:
5 August 2022

The CMC certification has been granted for the first time on:
5 August 2020

The CMC certification valid for the following test facilities:

XCaliber flow loop
VSL CMC certificate number 39333654



Manager Liquid Flow Metrology VSL, F.M. Smits