



VSL, the National Metrology Institute (NMI) of the Netherlands declares that the Calibration and Measurement Capabilities of the calibration test facilities named in this certificate are granted until the expiration date of their own CMC certificate to:

Thijsseweg 11
2629 JA Delft
The Netherlands
www.vsl.nl

LLC KROHNE-Automatika

Settlement Stromilovo | Volzhsky region 443532 Samara area | Russian Federation

During the CMC certification process it was demonstrated that the test facilities named in this certificate apply to the following requirements:

- All instruments in the test facilities are traceable to international accepted measurement standards,
- The calibration models (equations) are correct,
- The Calibration and Measurement Capabilities (CMC) are calculated according the JCGM 100:2008 "Evaluation of measurement data - Guide to the expression of uncertainty in measurement",
- A set of procedures is in place to maintain the CMC for the period of validation.

LF.2015.01.0010

granted on 21 November 2019


This declaration is valid until:
21 November 2021

The CMC certification has been granted for the first time on:

2 September 2015

Test facilities with CMC certification valid under this certificate:

**Flow meter calibration facility reference tower
PSTR04, VSL CMC certificate number
39333541, valid until 21 November 2021**



Manager Flow Metrology VSL, F.M. Smits