



VSL CMC Certified – Scope of facilities under certification

CMC Certification number	LF.2008.04.0002
Valid until	12 December 2020
Company	SENSIA
Address	1000 McClaren Woods Drive Coraopolis, PA 15108 United States of America

Scope of facilities under certification

Identification facility	Certified range	CMC	Test liquid	Remarks
Large liquid flow meter calibration loop	10 to 3900 m ³ /h	0.03 to 0.09%	Exxsol D80 Drakeol 5 Drakeol 32	CMC range depending on reference standard: pipe prover(s) or master meter(s).

