



## VSL CMC Certified – Scope of facilities under certification

<b>CMC Certification number</b>	<b>LF.2015.01.0010</b>
<b>Valid until</b>	21 November 2021
<b>Company</b>	LLC KROHNE-Automatika
<b>Address</b>	Settlement Stomilovo Volzhsky region 443532 Samara area Russian Federation

### Scope of facilities under certification

Identification facility	Certified range	CMC	Test liquid	Remarks
Liquid flow meter calibration facility: PSTR04	(0.007 – 2.8) m <sup>3</sup> /s	0.027% <sup>1)</sup>	Water	1) minimum reference volume 22 m <sup>3</sup> 2) minimum reference volume of 2.8 m <sup>3</sup>  Next to these criteria the criteria as set in VSL CMC Certificate 39333541 must be met.
	(0.0007 – 2.8) m <sup>3</sup> /s	0.092% <sup>2)</sup>		

